and purity. Ampicillin concentrated stock solutions for use in preparing antimicrobial susceptibility test panels are frozen aqueous ampicillin trihydrate stock solutions serially diluted with distilled water to contain approximately 3,200, 1,600, 800, 400, 200, 100, and 50 micrograms ampicillin per milliliter. The potency of each solution is satisfactory if it is not less than 100 percent and not more than 150 percent of the number of micrograms of ampicillin that it is represented to contain. The pH of the solution containing 3,200 micrograms of ampicillin per milliliter is not less than 4.0 and not more than 7.0. The ampicillin trihydrate used conthe requirements §440.7(a)(1) (i), (iii), (iv), (v), (vi), (vii), and (viii) of this chapter.

- (2) Labeling. It shall be labeled in accordance with the requirements of § 432.5 of this chapter.
- (3) Requests for certification; samples. In addition to complying with the requirements of §431.1 of this chapter, each such request shall contain:
- (i) Results of tests and assays on the batch for potency and pH.
- (ii) Samples required: A minimum of five frozen aliquots of each dilution of the concentrated stock solutions, each containing at least 2.5 milliliters.
- (b) *Tests and methods of assay.* The sample solutions must be thawed and brought to room temperature before testing.
- (1) Potency. Proceed as directed in §436.105 of this chapter, preparing the sample for assay as follows: Dilute an accurately measured representative portion of the sample with 0.1M potassium phosphate buffer, pH 8.0 (solution 3), to the reference concentration of 0.1 microgram of ampicillin per milliliter (estimated).
- (2) pH. Proceed as directed in §436.202 of this chapter, using the solution containing 3,200 micrograms of ampicillin per milliliter.

## § 460.113 Carbenicillin concentrated stock solutions for use in antimicrobial susceptibility test panels.

(a) Requirements for certification—(1) Standards of identity, strength, quality, and purity. Carbenicillin concentrated stock solutions for use in preparing

antimicrobial susceptibility test panels are frozen aqueous carbenicillin disodium stock solutions serially diluted with distilled water to contain approximately the following concentrations: 20,480, 10,240, 5,120, 2,560, 1,280, 640, and 320 micrograms of carbenicillin per milliliter. The potency of each solution is satisfactory if it is not less than 100 percent and not more than 150 percent of the number of micrograms of carbenicillin that it is represented to contain. The pH of the solution containing 20,480 micrograms of carbenicillin per milliliter is not less than 6.0 and not more than 8.0. The carbenicillin disodium used conforms to the requirements of §440.13a (a)(1) (i), (v), (vi), and (vii) of this chapter.

- (2) Labeling. It shall be labeled in accordance with the requirements of § 432.5 of this chapter.
- (3) Requests for certification; samples. In addition to complying with the requirements of §431.1 of this chapter, each such request shall contain:
- (i) Results of tests and assays on the batch for potency and pH.
- (ii) Samples required: A minimum of five frozen aliquots of each dilution of the concentrated stock solutions, each containing at least 5 milliliters.
- (b) Tests and methods of assay. The sample solutions must be thawed and brought to room temperature before testing.
- (1) Potency. Proceed as directed in §436.105 of this chapter, preparing the sample for assay as follows: Dilute an accurately measured representative portion of the sample with 1.0 percent potassium phosphate buffer, pH 6.0 (solution 1), to the reference concentration of 20 micrograms of carbenicillin per milliliter (estimated).
- (2) pH. Proceed as directed in §436.202 of this chapter, using the solution containing 20,480 micrograms of carbenicillin per milliliter.

## § 460.116 Cephalothin concentrated stock solutions for use in antimicrobial susceptibility test panels.

(a) Requirements for certification—(1) Standards of identity, strength, quality, and purity. Cephalothin concentrated stock solutions for use in preparing susceptibility test panels are frozen